

InQwire®

Simplicity at its Best

HYDROPHILIC MARKER WIRE

PLATINUM COIL TIP



TRANSITION



MARKER BANDS



*A wire featuring marker bands
on a hydrophilic coated, stainless
steel shaft with a platinum tip*

 MERT MEDICAL

CONVENIENCE, SMOOTHNESS, & CLARITY IS AS SIMPLE AS 1-2-3

Marks are placed at both 5cm & 10cm increments for increased accuracy. The marks are laser etched for a seamless transition between the marks and the wire body shaft. A black marker placed against the stainless steel wire body increases visibility.



Depth is determined from center to center of each band(s).

The First mark on the wire is at the 20cm length from the distal tip.

A 5cm "hash mark" is placed between each 10cm marker.

At every 10cm length there is a bold marker corresponding to the length measurement.



Smooth, seamless transition from wire body to tip.



20 cm platinum coil tip for increased visibility under fluoroscopy.

MARKER GUIDE WIRE INFORMATION:

Part # 6676-H1 Straight tip .018" (.046mm) x 80cm

Part # 6676-G1 Straight tip .018" (.046mm) x 120cm

Description: Stainless Steel Body Shaft, with Hydrophilic Coating
20cm Platinum Coil Straight Tip

Common Uses: PICC line placement and repositioning
Other catheter insertions such as Central venous lines (CVP) Swan Ganz placements/repositioning, etc...
Any similar type of catheter insertion procedure where a measurement tool may be needed.

Packaging: Sterile Single Pouch - Box of 10 Units.

 **MERITMEDICAL**
Innovations in Cardiology & Radiology

Merit Medical Systems, Inc. • 1600 West Merit Parkway, • South Jordan, Utah 84095 • 1-801-253-1600 • 1-800-35-MERIT
Merit Medical Europe • Horsterweg 24 • 6199 AC MAASTRICHT-AIRPORT • The Netherlands • +31 43 35 88 222
Merit Medical Ireland Ltd. • Parkmore Business Park West • Galway, Ireland • +353 (0) 91 703 733
Free phone from specific country:

Belgium 0800 72906 • France 0800 91 60 30 • Germany 0800 182 0871 • Luxembourg 0800 22522 • Netherlands 0800 022 8184 • UK 0800 973 115

www.merit.com

400711001/A ID 070805